



# Agilent 6495 Triple Quadrupole LC/MS with iFunnel Technology

## Data Sheet



The Agilent 6495 Triple Quadrupole LC/MS enables the lowest limits of detection and quantitation with performance specifications in Signal-to-Noise ratio (S/N) and Instrument Detection Limit (IDL). IDL is a rigorous, statistically based metric that indicates practical sensitivity performance of your quantitative assays. The Agilent 6495 Triple Quadrupole LC/MS achieves sensitivity and resolution specifications with autotune.

Parameter	Measure	Specification
MRM sensitivity S/N ESI positive	1 pg of reserpine injected on column, quantifying on the transition $m/z$ 609 to 195	S/N > 150,000:1 Noise 1 x RMS
MRM sensitivity S/N ESI negative	1 pg of chloramphenicol injected on column, quantifying on the transition $m/z$ 321 to 152	S/N > 150,000:1 Noise 1 x RMS
MRM sensitivity IDL ESI positive	1 fg of reserpine injected on column, quantifying on the transition $m/z$ 609 to 195	IDL < 0.75 fg
MRM sensitivity IDL ESI negative	1 fg of chloramphenicol injected on column, quantifying on the transition $m/z$ 321 to 152	IDL < 0.75 fg
Mass range		5 – 2,250 $m/z$
Polarity switching		30 ms
Mass resolution (autotune)	Full width at half maximum	0.7 Da
Mass resolution (manual tune)	Full width at half maximum	0.5 Da
Mass accuracy		0.1 Da from 5 – 1,000 $m/z$ 0.2 Da from 1,000 - 2,000 $m/z$ 0.3 Da from 2,000 - 2,250 $m/z$
Mass stability		< 0.1 Da in 24 hours
Dynamic range		> $6.0 \times 10^6$
Scan modes		MS scan, MS/MS product ion scan, MRM, MS/MS neutral loss/gain scan and precursor ion scan, SIM
Maximum scan rate		15,000 Da/s
Maximum MRM acquisition rate		500 MRMs/sec
Minimum MRM dwell time		1 ms
MRM transitions		450 per time segment, > 40,000 ion transitions per method
Dynamic MRM transitions		4,000 ion transitions per method
Triggered MRM transitions		Up to 10 MRM transitions (primary and secondary) for library search and compound confirmation
Collision cell ion clearance		< 1 ms



**Agilent Technologies**

## General system specifications

Parameter	Specification
Single point of control	Single-point data system method capability with full control of Agilent 1200 Series HPLC systems and 6495A Triple Quadrupole LC/MS System
Time programming	<ul style="list-style-type: none"><li>• Polarity change in time segment</li><li>• Scan and SIM or MRM (plus other modes of data collection)</li><li>• Dynamic and triggered MRM aligns MRMs with compound retention time</li><li>• Solvent divert through calibrant delivery system valve</li></ul>
Wide range of ionization sources	<ul style="list-style-type: none"><li>• Electrospray (ESI)</li><li>• Nanospray with HPLC-Chip Cube MS interface</li><li>• APCI source (Atmospheric Pressure Chemical Ionization)</li><li>• Multimode source (simultaneous ESI and APCI)</li><li>• APPI source (Atmospheric Pressure Photo Ionization)</li></ul>
Autotune	Automated optimization of ion optics and mass axis calibration in positive and negative ion modes using a proprietary tune solution
Solvent declustering	Countercurrent gas
Detector	High-energy conversion dynode and high-gain electron multiplier horn
Vacuum system	Two turbomolecular pumps with two mechanical pumps

## Ordering information

### G6495AA: 6495 Triple Quadrupole LC/MS System

Includes the 6495 Triple Quadrupole Mass Spectrometer, MassHunter Workstation Software with both compliance and method optimization software, a PC, a monitor and printer, and service installation of the system.

The above are not standard installation specifications for the 6495 Triple Quad. Performance specifications in this document are reviewed for accuracy, but they do not represent the tests and procedures performed at installation, which are described in the Agilent 6400 Series Triple Quad LC/MS System Installation Manual, document G3335-90170 or subsequent version number. See Site Preparation Guide and Service Notes for additional product and specification information.

[www.agilent.com/chem/qqq](http://www.agilent.com/chem/qqq)

Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc. 2014  
Published in the U.S.A. June 5, 2014  
5991-4704EN



**Agilent Technologies**